

2085 | 备用测试盖：7103



产品概述: 2085 | 备用测试盖：7103

This spare test lid is for the 7103 Micro-Bath, because, sometimes you need a lid with holes drilled in it and sometimes you need a lid that is completely sealed. For calibration, it is convenient to drill a pattern of holes into a test lid to keep the bath sealed during calibration for optimal stability, and to ensure probes are properly spaced from one another and go into the bath vertically. However, you also want to have a lid that prevents spilling when you are carrying the bath from the lab or shop to the test site, so we recommend keeping a spare lid to use during testing. That way you can have ideal test conditions and a clean work environment.

型号



2085

备用测试盖：7103



Fluke. 让您的工作畅通无阻。

福祿克測試儀器（上海）有限公司 電話：400-810-3435 ©2025 福祿克公司
07/2025
北京福祿克世祿儀器維修和服務有限公司 電
話：400-615-1563 未經許可，本文檔禁止修改
福祿克測試儀器（上海）有限公司上海維修中心 電
話：021-54402301, 021-54401908分機269
福祿克測試儀器（上海）有限公司深圳第一特約維修點
電話：0755-86337229